






REV.	LCO No.	REV DESCRIPTION	DATE	APPROVED
A	---	INITIAL RELEASE	-----	---



24" X 24" CLEAR ACRYLIC
VACUUM CHAMBER W/ HINGED
LID AND 2X LID CLAMPS

ALUMINUM FRAME W/ 2X FIXED AND
2X LOCKING SWIVEL CASTERS



- NOTES:
1. CHAMBER LEAK TESTED TO 1E-6 ATMCC/SEC.
 2. CHAMBER RATED FROM ATM TO 29.9 inHg OR 1 TORR.



UNLESS OTHERWISE SPECIFIED:
DIMENSIONS ARE IN INCHES.
TOLERANCES:
FRACTIONAL: ±1/16"
ANGULAR: MACH ±1° / BEND ±2°
TWO PLACE DECIMAL: ±.03
THREE PLACE DECIMAL: ±.005
GEOMETRIC TOLERANCING PER
ASME Y14.5M - 2009

PROPRIETARY AND CONFIDENTIAL
THE INFORMATION CONTAINED IN THIS
DRAWING IS THE SOLE PROPERTY OF LACO
TECHNOLOGIES, INC. ANY
REPRODUCTION IN PART OR AS A WHOLE
WITHOUT THE WRITTEN PERMISSION OF
LACO TECHNOLOGIES, INC., IS PROHIBITED.

**ASSY, TOP LEVEL, 24X24
ACRYLIC PACKAGE TESTER
SYSTEM WITH PUMP, STANDARD**

DRAWN BY	NAME	DATE
JRA	JRA	30 JAN 2019

SEE ABAS FOR APPROVER	DWG. NO.	REV
	115533-C	A
SCALE: 1:16	SIZE: B	SHEET 1 OF 1