



EXTERNAL HELIUM CALIBRATED LEAK STANDARD

Fast • Accurate • Accredited

Part Number: CM511.0-44121VG/2

Leak Rate: 1.0×10^{-4} std cc/sec
Leak Element: Micro-Tube Capillary
Gas: Helium (100%)
of Cal Pts: 4
Reservoir Style: 300 cc DOT
Est. Depl. Rate: 11%/year
Isolation Valve: Isolation Valve, Manual
Outlet Connection: NW25 Flange
Outlet Pressure: Vacuum
Additional Options: Pressure Gauge

Includes calibration that is NIST
traceable and A2LA accredited



Calibration Cert. No. 1530.01

PRESSURE GAUGE

SS 300CC
RESERVOIR

MANUAL ISO
VALVE

NW25 FLANGE

